

- [contributing](#)
- [courses](#)
- [epub](#)
- [faq](#)
 - [old_dongles_drivers_for_fluofit_symphotime_32](#)
 - [how_can_i_analyze_my_data_using_non-picoquant_software](#)
 - [laser_intensity_response](#)
 - [novaflim_calculated_irf](#)
 - [novaflim_license](#)
 - [the_dark_counts_of_my_spads_increased_when_do_i_need_to_take_action](#)
 - [why_do_i_not_get_the_same_count_rates_on_both_spad_detectors](#)
- [fluotime250](#)
- [general](#)
 - [advantages_and_disadvantages_of_two_photon_excitation_tpe](#)
 - [community_news](#)
 - [fluorescence_correlation_spectroscopy-a_short_introduction](#)
- [glossary](#)
 - [alex](#)
 - [aotf](#)
 - [asymptotic_standard_errors](#)
 - [bdf](#)
 - [bifl](#)
 - [bootstrap_method](#)
 - [bootstrap](#)
 - [cfd](#)
 - [chi_square_management](#)
 - [chi_square](#)
 - [convolution](#)
 - [dead_time](#)
 - [deconvolution](#)
 - [differential_count_rate](#)
 - [fast_flim](#)
 - [fast_lifetime](#)
 - [fastlt](#)
 - [fcs](#)
 - [flim](#)
 - [fluorescence_lifetime](#)
 - [frap](#)
 - [fret](#)
 - [fwhm](#)
 - [hybrid_pmt](#)
 - [irf](#)
 - [least_squares](#)
 - [marquardt-levenberg](#)
 - [mcp](#)
 - [mcs](#)

- mle
- monte_carlo
- nim
- od
- pie
- pile-up_effect
- pmt
- poisson_distribution
- pre-histogrammed_image
- pulsed_interleaved_excitation
- reconvolution
- residuals
- spad
- spads
- support_plane_analysis
- t2-mode
- t3-mode
- tcspc
- ttr
- **howto**
 - ax_flim_example
 - ax_flim_examples
 - burst_duration_and_fcs_diffusion_time
 - phasor_plot_structure_separation
 - 2ffcs
 - align_beam_backreflection
 - antibunching_measurements
 - avoid_pile_up_effect_in_flim_measurements
 - calculate_and_fit_fcs_traces_with_the_fcs_script
 - calculate_fccs_trace_with_the_grouped_fcs_script
 - calculate_ratiometric_fret-images
 - calculate_ratiometric_single_pair_fret_distributions_using_the_pie-fret_script
 - calculate_ratiometric_single_pair_fret_distributions
 - calibrate_the_confocal_volume_for_fcs_using_the_fcs_calibration_script
 - check_overlap_of_different_color_confocal_volumes
 - data_file_import
 - determination_of_the_focal_width_with_the_focal
 - diamond_nv_centers
 - exchange_dichroic_mt200
 - flim_fret_calculation_for_multi_exponential_donors
 - flim_measurement_using_a_nikon_a1_with_a_flim_and_fcs_upgrade
 - flim-fret_calculation_for_single_exponential_donors
 - flim-fret_measurement_using_an_olympus_fv1200_with_a_flim_and_fcs_upgrade
 - how_to_measure_the_instrument_response_function_irf
 - how_to_work_with_the_instrument_response_function_irf
 - intensity_time_trace_analysis
 - lifetime_fitting_using_the_flim_analysis
 - lifetime_fitting_using_the_tcspc_fitting_script
 - lifetime-fitting_using_the_flim_script
 - lifetime-fitting_using_the_rapid_reconvolution_algorithm
 - measuring_quantum_yield
 - mt200everyday_alignment
 - mt200fcs
 - mt200fundamental_alignement
 - pattern_matching

- [performing_an_fcs_measurement_with_an_olympus_fv1200_upgrade_kit](#)
- [phasor_analysis](#)
- [reconvolution_fit](#)
- [recording_a_fluorescence_lifetime_image_flim_stack_with_a_lsm_upgrade_kit_on_a_nikon_a1](#)
- [registering_new_scripts](#)
- [roi_fitting_using_the_flim_script](#)
- [select_the_correct_pinhole_size](#)
- [separation_of_2_species_with](#)
- [symphotime_tips_and_tricks](#)
- [t3r_antibunching_-_slow_decay](#)
- [update](#)
- [using_the_anisotropy_image_script](#)
- [using_the_antibunching_script](#)
- [using_the_flcs_script_for_spectral_crosstalk_removal_via_flccs](#)
- [visualizing_dynamics_using_the_multiframe-flim_script](#)
- [visualizing_dynamics_with_the_multi_frame_flim_analysis](#)
- [playground](#)
 - [test](#)
- [products](#)
 - [hydrharp_400](#)
 - [microtime](#)
 - [picoharp_300](#)
 - [sepia_ii](#)
 - [symphotime64](#)
 - [tcspc_electronics](#)
- [software](#)
 - [easytau](#)
 - [fcs_viewer](#)
 - [flimfit](#)
 - [fluofit](#)
 - [pycorrfit](#)
 - [supported_mt200_pc_configuration_for_symphotime_32](#)
 - [supported_mt200_pc_configuration_for_symphotime_64](#)
 - [symphotime_32](#)
 - [symphotime](#)
 - [symphotime64](#)
- [support](#)
 - [configuring_symphotime64_after_installation](#)
 - [supported_mt200_pc_configuration_for_symphotime_32](#)
 - [supported_mt200_pc_configuration_for_symphotime_64](#)
 - [tcspc_external_markers](#)
- [tag](#)
 - [howto](#)
- [technical_docs](#)
 - [beampath_of_the_zeiss_lsm700](#)
- [wiki](#)
 - [dokuwiki](#)
 - [ebook](#)
 - [syntax](#)
 - [welcome](#)
- [writingroom](#)
- [applications](#)
- [basics](#)
- [beampath_of_the_zeiss_lsm880](#)
- [contributions](#)

- [contributions12](#)
- [create_time_gated_image](#)
- [data_analysis](#)
- [drafts](#)
- [flim_fcs_using_olympus_fluoview_fv3000_lsm_upgrade_kit](#)
- [fluorescence_lifetime_measurements_using_the_fluotime_300](#)
- [fluorophores_and_samples](#)
- [fullindex](#)
- [how_to_reinstall_symphotime](#)
- [imprint](#)
- [interfacing_time_resolved_spectrometer_fluotime_300_microscope_microtime_100](#)
- [laser_safety_instructions](#)
- [legal_information](#)
- [lifetime_component_decomposition_phasor_plot_analysis](#)
- [lsm710](#)
- [measurement_hardware_instrumentation](#)
- [privacy_policy](#)
- [some_origins_of_multiexponential_decays_for_single_dyes](#)
- [supported_th260_pc](#)
- [synchrotron_application](#)
- [synchrotron_applications](#)
- [tutorials](#)
- [video_tutorials](#)

Copyright of this document belongs to PicoQuant GmbH. No parts of it may be reproduced, translated or transferred to third parties without written permission of PicoQuant GmbH. All information given here is reliable to our best knowledge. However, no responsibility is assumed for possible inaccuracies or omissions. Specifications and external appearances are subject to change without notice.



PicoQuant GmbH
Rudower Chaussee 29 (IGZ)
12489 Berlin
Germany

P +49-(0)30-1208820-89
F +49-(0)30-1208820-90
info@picoquant.com
www.picoquant.com