

MT200 Fundamental Alignment

This procedure is a quick check of the optical hardware and evaluates the overall system performance. It takes typically ten to fifteen minutes and allows the overall alignment to be easily maintained. The basic idea is to make sure that the laser beam is going properly in the Main Optical Unit (MOU). To this end we follow the excitation beam path step by step from the Laser Combining Unit (LCU) to the major dichroic, then through the pinhole and finally to the detectors.

Copyright of this document belongs to PicoQuant GmbH. No parts of it may be reproduced, translated or transferred to third parties without written permission of PicoQuant GmbH. All information given here is reliable to our best knowledge. However, no responsibility is assumed for possible inaccuracies or omissions. Specifications and external appearances are subject to change without notice.



PicoQuant GmbH
Rudower Chaussee 29 (IGZ)
12489 Berlin
Germany

P +49-(0)30-1208820-89
F +49-(0)30-1208820-90
info@picoquant.com
www.picoquant.com