

## FLIM Measurement Using a Zeiss LSM710/LSM780/LSM880 with a FLIM and FCS Upgrade

This tutorial shows the recording of FLIM images using an LSM upgrade kit, in this case a Zeiss LSM710. In FLIM-measurements, the fluorescence lifetime of a component can be determined which can be used e.g.

- to distinguish different tissue or cell constituents
- to monitor concentration changes using a reporter dye
- To monitor interactions between two molecules via FRET.

The data acquisition is analogous also for upgrades on an LSM780 or LSM880, and the general principles of the data acquisition are analogous in all LSM upgrade kits.

Continue to

- [How to measure the Instrument Response function \(IRF\)](#)
- [FLIM FRET Calculation for Multi Exponential Donors](#)
- [FLIM ROI Fitting](#)
- [FLIM-FRET Calculation for Single Exponential Donors](#)
- [Lifetime Fitting Using the FLIM Analysis](#)
- [Pattern Matching](#)
- [Phasor Analysis](#)
- [Visualizing Dynamics with the Multi Frame FLIM Analysis](#)

Copyright of this document belongs to PicoQuant GmbH. No parts of it may be reproduced, translated or transferred to third parties without written permission of PicoQuant GmbH. All information given here is reliable to our best knowledge. However, no responsibility is assumed for possible inaccuracies or omissions. Specifications and external appearances are subject to change without notice.



PicoQuant GmbH  
Rudower Chaussee 29 (IGZ)  
12489 Berlin  
Germany

P +49-(0)30-1208820-89  
F +49-(0)30-1208820-90  
info@picoquant.com  
www.picoquant.com