

Synchrotron Applications

TCSPC devices (such as [PicoHarp 300](#)) are also used to measure the temporal distribution of synchrotron radiation that corresponds to the temporal distribution of the electron beam that generate it and thus the filling pattern. We measure the temporal distribution of the radiation by measuring the arrival time of a photon to the laboratory with respect to a trigger given by the machine timing. So in this optics we are certainly measuring the time of flight of photons.

Software to integrate the [PicoHarp 300](#) into the [Tango Control Software](#) used at the [ALBA Synchrotron](#) is being developed.

The [development](#) has been in split into smaller parts:

- A cython library to access the original's manufacturer library from python and in a pythonic way,
- An agent in the [Tango distributed control system](#) that manage the instrument,
- A [Taurus gui](#) with access to the instrument from the control system

There is a development now for another agent in the DCS to process the histogram given by the PicoHarp300 instrument to have the distribution mention before.

Please check <http://www.picoquant.com/applications/category/metrology/bunch-purity-measurements> for other Synchrotron-related applications.

Copyright of this document belongs to PicoQuant GmbH. No parts of it may be reproduced, translated or transferred to third parties without written permission of PicoQuant GmbH. All information given here is reliable to our best knowledge. However, no responsibility is assumed for possible inaccuracies or omissions. Specifications and external appearances are subject to change without notice.



PicoQuant GmbH
Rudower Chaussee 29 (IGZ)
12489 Berlin
Germany

P +49-(0)30-1208820-89
F +49-(0)30-1208820-90
info@picoquant.com
www.picoquant.com